

Translation

10/541297
PATENT COOPERATION TREATY

PCT/DE2003/004017



PCT

INTERNATIONAL PRELIMINARY EXAMINATION REPORT

(PCT Article 36 and Rule 70)

Applicant's or agent's file reference PCT 7054	FOR FURTHER ACTION See Notification of Transmittal of International Preliminary Examination Report (Form PCT/IPEA/416)	
International application No. PCT/DE2003/004017	International filing date (day/month/year) 05 December 2003 (05.12.2003)	Priority date (day/month/year) 03 January 2003 (03.01.2003)
International Patent Classification (IPC) or national classification and IPC A61B 1/00		
Applicant KRESS, Jürgen		

1. This international preliminary examination report has been prepared by this International Preliminary Examining Authority and is transmitted to the applicant according to Article 36.

2. This REPORT consists of a total of 4 sheets, including this cover sheet.

☒ This report is also accompanied by ANNEXES, i.e., sheets of the description, claims and/or drawings which have been amended and are the basis for this report and/or sheets containing rectifications made before this Authority (see Rule 70.16 and Section 607 of the Administrative Instructions under the PCT).

These annexes consist of a total of 5 sheets.

3. This report contains indications relating to the following items:

- I ☒ Basis of the report
- II ☐ Priority
- III ☐ Non-establishment of opinion with regard to novelty, inventive step and industrial applicability
- IV ☐ Lack of unity of invention
- V ☒ Reasoned statement under Article 35(2) with regard to novelty, inventive step or industrial applicability; citations and explanations supporting such statement
- VI ☐ Certain documents cited
- VII ☐ Certain defects in the international application
- VIII ☐ Certain observations on the international application

Date of submission of the demand 14 April 2004 (14.04.2004)	Date of completion of this report 01 October 2004 (01.10.2004)
Name and mailing address of the IPEA/EP	Authorized officer
Facsimile No.	Telephone No.

INTERNATIONAL PRELIMINARY EXAMINATION REPORT

International application No.

PCT/DE2003/004017

I. Basis of the report

1. With regard to the elements of the international application:*

- ☐ the international application as originally filed
- ☒ the description:
 pages 2-12, as originally filed
 pages _____, filed with the demand
 pages 1, 1a, 3a, 3b, filed with the letter of 21 September 2004 (21.09.2004)
- ☒ the claims:
 pages 2-28, as originally filed
 pages _____, as amended (together with any statement under Article 19
 pages _____, filed with the demand
 pages 1, filed with the letter of 21 September 2004 (21.09.2004)
- ☒ the drawings:
 pages 1/3-3/3, as originally filed
 pages _____, filed with the demand
 pages _____, filed with the letter of _____
- ☐ the sequence listing part of the description:
 pages _____, as originally filed
 pages _____, filed with the demand
 pages _____, filed with the letter of _____

2. With regard to the language, all the elements marked above were available or furnished to this Authority in the language in which the international application was filed, unless otherwise indicated under this item.
 These elements were available or furnished to this Authority in the following language _____ which is:
- ☐ the language of a translation furnished for the purposes of international search (under Rule 23.1(b)).
- ☐ the language of publication of the international application (under Rule 48.3(b)).
- ☐ the language of the translation furnished for the purposes of international preliminary examination (under Rule 55.2 and/or 55.3).

3. With regard to any nucleotide and/or amino acid sequence disclosed in the international application, the international preliminary examination was carried out on the basis of the sequence listing:

- ☐ contained in the international application in written form.
- ☐ filed together with the international application in computer readable form.
- ☐ furnished subsequently to this Authority in written form.
- ☐ furnished subsequently to this Authority in computer readable form.
- ☐ The statement that the subsequently furnished written sequence listing does not go beyond the disclosure in the international application as filed has been furnished.
- ☐ The statement that the information recorded in computer readable form is identical to the written sequence listing has been furnished.

4. ☐ The amendments have resulted in the cancellation of:

- ☐ the description, pages _____
- ☐ the claims, Nos. _____
- ☐ the drawings, sheets/fig _____

5. ☐ This report has been established as if (some of) the amendments had not been made, since they have been considered to go beyond the disclosure as filed, as indicated in the Supplemental Box (Rule 70.2(c)).**

* Replacement sheets which have been furnished to the receiving Office in response to an invitation under Article 14 are referred to in this report as "originally filed" and are not annexed to this report since they do not contain amendments (Rule 70.16 and 70.17).

** Any replacement sheet containing such amendments must be referred to under item 1 and annexed to this report.

INTERNATIONAL PRELIMINARY EXAMINATION REPORT

International application No.

PCT/DE 03/04017

V. Reasoned statement under Article 35(2) with regard to novelty, inventive step or industrial applicability; citations and explanations supporting such statement

1. Statement

Novelty (N)	Claims	1-28	YES
	Claims		NO
Inventive step (IS)	Claims	1-28	YES
	Claims		NO
Industrial applicability (IA)	Claims	1-28	YES
	Claims		NO

2. Citations and explanations

- US-A-4 646 722 (D1), which is acknowledged by the applicant in the description, discloses an endoscope with a hygienic protective device according to the preamble of independent claim 1.

The problem to be solved by the characterizing features of the present application can be regarded as that of quickly producing a secure connection between the cover and the endoscope by applying negative pressure between the cover and the endoscope shaft to suction away the air present there.

Another document, EP-A-O 310 515, discloses the application of negative pressure between a dental camera and a protective device for the camera via channels integrated into the camera. However, a combination of features according to claim 1 of the present application can also be used irrespective of the type of endoscope owing to the integration of the channels into the endoscope protective device and is therefore neither known nor derivable in an obvious way from D1 or the other remaining documents. Claim 1 therefore meets the requirements of PCT Article 33(2) to (4).

INTERNATIONAL PRELIMINARY EXAMINATION REPORT

International application No.

PCT/DE 03/04017

2. Independent claims 27 and 28 describe the use of the claimed device with the characterizing features and therefore also meet the PCT requirements for novelty and inventive step. The same applies to claims 2 to 26, which are dependent on claim 1.